

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000919  
10:41 10/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:53  
PRESSURIZING 3:00  
STERILIZING 4:07  
136.4 C 232kPa 4:07  
136.4 C 233kPa 4:37  
135.9 C 230kPa 5:07  
136.3 C 231kPa 5:37  
136.4 C 233kPa 6:07  
136.1 C 230kPa 6:37  
136.2 C 231kPa 7:07  
136.4 C 233kPa 7:37  
136.4 C 233kPa 7:37  
VENTING 7:37  
AIR DRYING 8:28  
CYCLE COMPLETE 18:39

-----  
Digital Signature #  
4D460660E95F72B8  
-----