
Midmark M11 - v1.0.5
Total Cycles: 2454

Name: RITTER2
Sterilizer ID: V2313228
06/01/2023 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.7 F 0.0
2:00 164.8 F 3.3
4:00 177.3 F 7.5
6:00 201.2 F 12.9
8:00 222.3 F 7.9
10:00 236.3 F 12.7
12:00 238.3 F 11.2
14:00 237.7 F 9.9
16:00 251.8 F 16.8
18:00 263.6 F 23.8
20:00 271.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 271.9 F 29.4
2:00 272.5 F 29.6
3:00 272.5 F 29.9
4:00 272.0 F 30.0

Min 271.5 F 29.3
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 217.4 F 0.0
10:00 225.8 F 0.0
15:00 215.5 F 0.0
20:00 226.1 F 0.0
25:00 225.1 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

06/01/2023 02:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____