
Midmark M11 - v1.0.0
Total Cycles: 2399

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2022 03:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 200.2 F 3.7
4:00 210.5 F 8.0
6:00 223.7 F 11.5
8:00 231.9 F 8.7
10:00 241.6 F 10.7
12:00 243.0 F 12.1
14:00 241.4 F 10.9
16:00 255.1 F 17.9
18:00 266.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.6
2:00 272.2 F 28.9
3:00 271.8 F 28.3
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 233.7 F 0.0
10:00 229.2 F 0.0
15:00 229.6 F 0.0
20:00 227.6 F 0.0
25:00 224.9 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

10/04/2022 04:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____