

-----  
Midmark M11 - v1.0.5  
Total Cycles: 928

Name: RITTER2  
Sterilizer ID: V2313228  
03/30/2022 10:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
2:00 184.0 F 0.4  
4:00 201.3 F 4.9  
6:00 215.8 F 10.4  
8:00 231.4 F 7.1  
10:00 241.2 F 13.1  
12:00 241.9 F 11.6  
14:00 241.2 F 10.9  
16:00 254.9 F 17.9  
18:00 266.6 F 25.1  
20:00 273.2 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.7 F 29.4  
2:00 272.5 F 29.8  
3:00 271.9 F 29.0  
4:00 272.9 F 29.6

Min 271.6 F 29.0  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 232.8 F 0.0  
10:00 230.1 F 0.0  
15:00 229.0 F 0.0  
20:00 226.4 F 0.0  
25:00 228.6 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

03/30/2022 11:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_