
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000225
8:48 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:19
PRESSURIZING 4:19
STERILIZING 5:52
136.0 C 223kPa 5:52
136.3 C 232kPa 6:22
136.2 C 231kPa 6:52
136.3 C 232kPa 7:22
136.3 C 232kPa 7:52
136.4 C 233kPa 8:22
136.4 C 233kPa 8:52
136.4 C 233kPa 9:22
136.4 C 233kPa 9:22
VENTING 9:22
AIR DRYING 10:26
CYCLE COMPLETE 25:26

Digital Signature #
3AF826C4711DD7B3
