
Midmark M11 - v1.0.0
Total Cycles: 2853

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2023 05:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.4 F 0.0
2:00 210.8 F 4.8
4:00 214.1 F 10.4
6:00 222.3 F 5.5
8:00 234.7 F 11.7
10:00 234.8 F 9.3
12:00 237.8 F 9.1
14:00 248.7 F 14.8
16:00 261.1 F 21.8
18:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.6
3:00 272.6 F 29.3
4:00 271.9 F 28.7

Min 271.5 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 223.1 F 0.0
10:00 225.8 F 0.0
15:00 226.4 F 0.0
20:00 228.1 F 0.0
25:00 229.4 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 18:29
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

02/02/2023 06:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____