
Midmark M11 - v1.0.0
Total Cycles: 4640

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2024 06:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 183.5 F 4.4
4:00 208.1 F 10.0
6:00 219.9 F 5.7
8:00 235.4 F 11.9
10:00 236.0 F 9.5
12:00 236.4 F 8.5
14:00 249.9 F 15.3
16:00 263.1 F 22.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.1 F 28.6
2:00 272.7 F 29.2
3:00 271.6 F 28.4
4:00 272.5 F 28.9

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 207.4 F 0.0
10:00 224.5 F 0.0
15:00 224.5 F 0.0
20:00 223.4 F 0.0
25:00 221.5 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

06/13/2024 07:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____