
Midmark M11 - v1.0.0
Total Cycles: 948

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2021 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 183.3 F 3.3
4:00 197.6 F 7.5
6:00 227.4 F 12.6
8:00 229.0 F 7.7
10:00 241.2 F 13.0
12:00 238.7 F 10.5
14:00 236.9 F 8.5
16:00 250.0 F 15.2
18:00 261.6 F 21.7
20:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.8
2:00 271.4 F 28.1
3:00 272.4 F 28.9
4:00 272.1 F 28.4

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 236.6 F 0.0
10:00 239.1 F 0.0
15:00 227.2 F 0.0
20:00 222.1 F 0.0
25:00 216.2 F 0.0
30:00 194.7 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

06/28/2021 01:18 PM
ACCEPTED: X
REJECTED:
Operator: U01