
Midmark M11 - v1.0.5
Total Cycles: 995

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.0 F 0.0
2:00 146.5 F 0.0
4:00 154.7 F 0.3
6:00 186.5 F 1.4
8:00 205.2 F 6.0
10:00 220.1 F 10.8
12:00 226.8 F 6.5
14:00 238.4 F 12.3
16:00 238.7 F 10.5
18:00 237.6 F 9.5
20:00 251.8 F 16.5
22:00 263.6 F 23.4
24:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.2
2:00 271.9 F 29.2
3:00 272.3 F 28.9
4:00 272.2 F 29.5

Min 271.5 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 197.2 F 0.0
10:00 203.2 F 0.0
15:00 206.8 F 0.0
20:00 206.9 F 0.0
25:00 210.2 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

04/18/2022 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____