
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001071
14:17 11/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:29
PRESSURIZING 4:30
STERILIZING 5:43
136.0 C 220kPa 5:43
136.3 C 223kPa 6:13
136.3 C 224kPa 6:43
136.0 C 223kPa 7:13
136.3 C 223kPa 7:43
136.4 C 224kPa 8:13
136.3 C 222kPa 8:43
136.4 C 225kPa 9:13
136.4 C 225kPa 9:13
VENTING 9:13
AIR DRYING 10:08
CYCLE COMPLETE 25:08

Digital Signature #
B661851221F6E699
