
Midmark M11 - v1.0.0
Total Cycles: 1915

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 154.2 F 2.3
4:00 176.3 F 6.0
6:00 197.5 F 10.6
8:00 220.6 F 4.3
10:00 231.7 F 9.5
12:00 234.8 F 9.2
14:00 242.1 F 11.6
16:00 238.9 F 9.7
18:00 252.6 F 16.4
20:00 264.3 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.5 F 28.8
2:00 271.5 F 28.0
3:00 271.8 F 28.6
4:00 272.6 F 28.6

Min 271.0 F 27.9
Max 273.1 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 232.8 F 0.0
10:00 156.1 F 0.0
15:00 142.9 F 0.0
20:00 141.0 F 0.0
25:00 137.4 F 0.0
30:00 132.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

05/05/2022 01:45 PM
ACCEPTED: X
REJECTED:
Operator: U02