
Midmark M11 - v1.0.0
Total Cycles: 4711

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2024 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.7 F 0.0
2:00 152.5 F 2.4
4:00 159.9 F 5.6
6:00 185.1 F 9.3
8:00 207.1 F 13.4
10:00 217.4 F 5.8
12:00 229.3 F 10.1
14:00 233.7 F 8.9
16:00 238.9 F 11.3
18:00 236.7 F 9.1
20:00 249.3 F 15.2
22:00 260.3 F 21.3
24:00 270.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.6 F 29.4
2:00 272.1 F 29.4
3:00 272.3 F 29.2
4:00 272.4 F 29.6

Min 271.7 F 28.9
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 203.7 F 0.0
10:00 204.9 F 0.0
15:00 203.9 F 0.0
20:00 215.2 F 0.0
25:00 230.9 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:38
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

07/10/2024 03:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____