
Midmark M11 - v1.0.0
Total Cycles: 4832

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2024 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.7 F 0.0
2:00 145.4 F 0.9
4:00 164.6 F 1.2
6:00 188.9 F 4.3
8:00 212.2 F 8.0
10:00 218.8 F 12.0
12:00 226.5 F 5.8
14:00 233.5 F 10.4
16:00 232.9 F 8.1
18:00 243.4 F 12.6
20:00 241.7 F 11.1
22:00 253.8 F 17.2
24:00 264.3 F 23.3
26:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.9
2:00 272.4 F 28.8
3:00 271.7 F 28.3
4:00 271.2 F 28.3

Min 271.2 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 240.7 F 0.0
10:00 220.5 F 0.0
15:00 231.5 F 0.0
20:00 198.0 F 0.0
25:00 176.3 F 0.0
30:00 173.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:16
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:55
CYCLE COMPLETE

08/19/2024 04:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____