
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001616
14:51 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:13
136.4 C 225kPa 4:13
136.4 C 226kPa 4:43
136.4 C 226kPa 5:13
136.1 C 223kPa 5:43
136.3 C 225kPa 6:13
136.4 C 226kPa 6:43
136.4 C 225kPa 7:13
136.2 C 224kPa 7:43
136.2 C 224kPa 7:43
VENTING 7:43
AIR DRYING 8:38
CYCLE COMPLETE 8:59

Digital Signature #
21C1FE6A7D6C9358
