
Midmark M11 - v1.0.5
Total Cycles: 2933

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.6 F 0.0
2:00 169.5 F 3.2
4:00 192.3 F 7.9
6:00 209.8 F 12.7
8:00 221.8 F 7.9
10:00 233.2 F 13.2
12:00 237.4 F 10.9
14:00 237.3 F 9.9
16:00 251.2 F 16.6
18:00 262.3 F 23.0
20:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.8
2:00 271.6 F 29.1
3:00 271.7 F 29.7
4:00 272.1 F 29.7

Min 271.1 F 29.0
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 212.8 F 0.0
10:00 221.0 F 0.0
15:00 213.5 F 0.0
20:00 208.7 F 0.0
25:00 207.6 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

09/26/2023 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____