
Midmark M11 - v1.0.5
Total Cycles: 3483

Name: RITTER2
Sterilizer ID: V2313228
03/27/2024 02:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.2 F 0.0
2:00 140.1 F 0.6
4:00 160.7 F 3.8
6:00 183.5 F 7.9
8:00 205.0 F 12.4
10:00 218.1 F 6.6
12:00 232.1 F 11.4
14:00 234.0 F 8.9
16:00 239.2 F 10.5
18:00 246.6 F 14.2
20:00 259.0 F 20.8
22:00 270.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.4
1:00 272.4 F 30.2
2:00 272.2 F 29.7
3:00 271.6 F 29.7
4:00 273.1 F 30.1

Min 271.4 F 29.3
Max 273.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.2 F 0.0
5:00 218.9 F 0.0
10:00 193.9 F 0.0
15:00 187.3 F 0.0
20:00 184.3 F 0.0
25:00 183.1 F 0.0
30:00 181.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

03/27/2024 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____