
Midmark M11 - v1.0.0
Total Cycles: 4554

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2024 08:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 71.3 F 0.0
4:00 86.1 F 0.2
6:00 115.9 F 1.7
8:00 145.0 F 4.3
10:00 165.3 F 7.4
12:00 184.3 F 11.0
14:00 216.8 F 4.0
16:00 219.1 F 5.9
18:00 228.5 F 9.4
20:00 234.8 F 13.3
22:00 236.3 F 9.3
24:00 240.9 F 10.5
26:00 244.2 F 12.3
28:00 255.5 F 18.1
30:00 265.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.6 F 28.9
2:00 270.9 F 28.1
3:00 272.6 F 29.2
4:00 271.5 F 28.5

Min 270.9 F 28.1
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 224.6 F 0.0
10:00 225.4 F 0.0
15:00 221.5 F 0.0
20:00 212.2 F 0.0
25:00 214.9 F 0.0
30:00 206.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:10
CYCLE COMPLETE

05/15/2024 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____