

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005514  
7:39 11/05/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:13  
STERILIZING 4:21  
136.4 C 220kPa 4:21  
136.4 C 220kPa 4:51  
136.3 C 222kPa 5:21  
136.4 C 220kPa 5:51  
136.3 C 222kPa 6:21  
136.4 C 222kPa 6:51  
136.4 C 223kPa 7:21  
136.5 C 221kPa 7:51  
136.5 C 221kPa 7:51  
VENTING 7:51  
AIR DRYING 8:41  
CASSETTE REMOVED  
CYCLE COMPLETE 22:46  
-----

Digital Signature #  
7F943BCA9A2D3039  
-----