
Midmark M11 - v1.0.5
Total Cycles: 1343

Name: RITTER2
Sterilizer ID: V2313228
07/26/2022 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 77.1 F 0.0
4:00 78.3 F 0.0
6:00 104.6 F 0.1
8:00 139.2 F 0.6
10:00 182.3 F 1.6
12:00 202.8 F 4.8
14:00 217.5 F 8.2
16:00 228.7 F 11.9
18:00 231.0 F 7.0
20:00 240.8 F 11.9
22:00 237.3 F 9.4
24:00 240.1 F 10.1
26:00 247.8 F 14.2
28:00 259.3 F 20.6
30:00 269.7 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.1 F 28.9
2:00 272.7 F 29.2
3:00 272.7 F 29.5
4:00 271.8 F 29.3

Min 271.4 F 28.8
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 233.4 F 0.0
10:00 213.1 F 0.0
15:00 205.6 F 0.0
20:00 208.1 F 0.0
25:00 230.2 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 31:40
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:55
CYCLE COMPLETE

07/26/2022 07:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____