
Midmark M11 - v1.0.0
Total Cycles: 1761

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2022 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 185.8 F 3.0
4:00 192.3 F 7.0
6:00 218.4 F 12.8
8:00 226.8 F 7.8
10:00 238.0 F 13.2
12:00 238.3 F 10.5
14:00 236.8 F 8.4
16:00 250.4 F 15.6
18:00 262.5 F 22.6
20:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.2
2:00 271.6 F 28.4
3:00 272.1 F 29.1
4:00 272.4 F 29.0

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 227.1 F 0.0
10:00 236.8 F 0.0
15:00 231.5 F 0.0
20:00 227.1 F 0.0
25:00 230.0 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

03/29/2022 12:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____