
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001507
9:00 10/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:19
PRESSURIZING 7:39
STERILIZING 9:22
136.0 C 220kPa 9:22
135.6 C 215kPa 9:52
136.2 C 220kPa 10:23
136.2 C 220kPa 10:53
136.4 C 219kPa 11:23
136.4 C 222kPa 11:53
136.3 C 221kPa 12:23
136.4 C 221kPa 12:53
VENTING 12:53
AIR DRYING 14:00
CYCLE COMPLETE 44:00

Digital Signature #
0B92D90FC6537500
