
Midmark M11 - v1.0.0
Total Cycles: 5388

Name: RITTER 1
Sterilizer ID: V2178198
03/05/2025 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.9 F 0.0
2:00 184.4 F 3.2
4:00 188.4 F 7.3
6:00 197.8 F 12.2
8:00 215.3 F 5.0
10:00 228.2 F 9.5
12:00 239.2 F 10.5
14:00 239.9 F 11.1
16:00 236.9 F 9.2
18:00 250.1 F 15.5
20:00 261.2 F 21.9
22:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.6 F 29.6
2:00 271.7 F 29.0
3:00 272.6 F 29.7
4:00 272.3 F 29.2

Min 271.7 F 29.0
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 227.9 F 0.0
10:00 223.0 F 0.0
15:00 211.2 F 0.0
20:00 209.4 F 0.0
25:00 199.2 F 0.0
30:00 200.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:27
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

03/05/2025 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____