
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001032
13:16 08/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 3:56
136.5 C 225kPa 3:56
136.4 C 225kPa 4:26
136.4 C 225kPa 4:56
136.4 C 225kPa 5:26
136.4 C 225kPa 5:56
136.2 C 224kPa 6:26
136.0 C 224kPa 6:56
136.3 C 223kPa 7:26
136.3 C 223kPa 7:26
VENTING 7:26
AIR DRYING 8:16
CYCLE COMPLETE 20:06

Digital Signature #
92FBFB1DD15AC381
