
Midmark M11 - v1.0.0
Total Cycles: 5245

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2025 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 165.4 F 0.8
4:00 170.3 F 4.7
6:00 199.2 F 8.9
8:00 218.5 F 6.4
10:00 224.4 F 8.2
12:00 236.9 F 13.1
14:00 237.1 F 9.7
16:00 237.2 F 8.6
18:00 248.0 F 14.0
20:00 259.9 F 20.5
22:00 270.6 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.2 F 28.7
2:00 271.9 F 28.3
3:00 271.8 F 28.5
4:00 272.4 F 28.6

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 201.4 F 0.0
10:00 199.6 F 0.0
15:00 229.2 F 0.0
20:00 231.2 F 0.0
25:00 230.5 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

01/15/2025 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____