

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2406

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2022 04:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
2:00 190.7 F 3.6  
4:00 207.7 F 7.3  
6:00 213.7 F 12.9  
8:00 227.7 F 7.3  
10:00 239.9 F 12.4  
12:00 237.7 F 9.8  
14:00 238.5 F 8.6  
16:00 249.0 F 14.8  
18:00 261.1 F 21.6  
20:00 272.2 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.0  
2:00 271.5 F 28.3  
3:00 272.5 F 29.2  
4:00 271.9 F 28.6

Min 271.3 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 239.5 F 0.0  
10:00 233.6 F 0.0  
15:00 229.8 F 0.0  
20:00 227.4 F 0.0  
25:00 226.7 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:31  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

10/05/2022 05:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_