
Midmark M11 - v1.0.5
Total Cycles: 1693

Name: RITTER2
Sterilizer ID: V2313228
11/07/2022 07:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 188.3 F 0.6
4:00 209.8 F 5.0
6:00 220.2 F 10.5
8:00 227.4 F 6.7
10:00 239.7 F 13.2
12:00 241.4 F 12.0
14:00 241.9 F 11.6
16:00 256.0 F 18.9
18:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.3
2:00 271.4 F 28.9
3:00 272.7 F 29.3
4:00 271.4 F 28.6

Min 271.3 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 223.9 F 0.0
10:00 229.0 F 0.0
15:00 228.5 F 0.0
20:00 228.2 F 0.0
25:00 228.5 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

11/07/2022 07:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____