

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001482  
8:37 10/07/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:04  
PRESSURIZING 6:55  
STERILIZING 8:44  
136.0 C 220kPa 8:44  
136.4 C 223kPa 9:14  
136.1 C 220kPa 9:44  
136.5 C 223kPa 10:14  
136.4 C 219kPa 10:44  
136.2 C 221kPa 11:14  
136.4 C 223kPa 11:44  
136.4 C 221kPa 12:14  
VENTING 12:14  
AIR DRYING 13:20  
CYCLE COMPLETE 43:20

-----  
Digital Signature #  
69A8D8A03520C0BA  
-----