
Midmark M11 - v1.0.0
Total Cycles: 5153

Name: RITTER 1
Sterilizer ID: V2178198
12/03/2024 04:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 168.8 F 0.2
4:00 177.0 F 0.6
6:00 195.6 F 3.1
8:00 214.3 F 7.0
10:00 226.0 F 11.2
12:00 227.6 F 6.4
14:00 240.4 F 11.9
16:00 238.1 F 9.8
18:00 238.5 F 9.1
20:00 249.7 F 14.9
22:00 261.4 F 21.4
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.1 F 28.6
2:00 272.6 F 28.7
3:00 272.2 F 28.3
4:00 271.5 F 27.9

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 226.9 F 0.0
10:00 218.1 F 0.0
15:00 246.1 F 0.0
20:00 250.6 F 0.0
25:00 239.5 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:03:24
CYCLE COMPLETE

12/03/2024 05:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____