

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4198

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2024 11:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.3 F 0.0  
2:00 155.3 F 1.6  
4:00 157.6 F 4.3  
6:00 182.3 F 7.8  
8:00 197.8 F 12.1  
10:00 212.7 F 4.5  
12:00 222.7 F 8.5  
14:00 233.0 F 12.9  
16:00 234.3 F 9.2  
18:00 240.7 F 10.6  
20:00 244.7 F 12.8  
22:00 255.9 F 18.6  
24:00 266.2 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.7 F 29.1  
2:00 271.4 F 28.5  
3:00 272.1 F 29.1  
4:00 272.4 F 28.9

Min 271.1 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 203.4 F 0.0  
10:00 211.7 F 0.0  
15:00 218.6 F 0.0  
20:00 210.3 F 0.0  
25:00 207.7 F 0.0  
30:00 206.3 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:46  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:04:33  
CYCLE COMPLETE

01/12/2024 12:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_