
Midmark M11 - v1.0.0
Total Cycles: 5132

Name: RITTER 1
Sterilizer ID: V2178198
11/26/2024 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.4 F 0.0
2:00 113.3 F 0.0
4:00 137.3 F 0.1
6:00 167.8 F 0.3
8:00 191.6 F 0.7
10:00 208.2 F 3.0
12:00 220.6 F 6.4
14:00 231.3 F 10.4
16:00 232.9 F 8.2
18:00 241.2 F 11.7
20:00 238.5 F 9.8
22:00 237.7 F 8.9
24:00 250.4 F 15.3
26:00 261.8 F 21.8
28:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.1 F 28.6
2:00 272.6 F 28.7
3:00 271.8 F 28.1
4:00 272.2 F 28.7

Min 271.6 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 184.0 F 0.0
10:00 193.0 F 0.0
15:00 197.5 F 0.0
20:00 204.8 F 0.0
25:00 226.3 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 28:25
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:53
CYCLE COMPLETE

11/26/2024 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____