
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001535
16:26 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:31
136.4 C 224kPa 3:31
136.3 C 226kPa 4:01
136.4 C 226kPa 4:31
136.3 C 225kPa 5:01
136.3 C 226kPa 5:31
136.4 C 226kPa 6:01
136.4 C 227kPa 6:31
136.3 C 225kPa 7:01
136.3 C 225kPa 7:01
VENTING 7:01
AIR DRYING 7:57
CYCLE COMPLETE 8:01

Digital Signature #
4D59A9F193738463
