
Midmark M11 - v1.0.5
Total Cycles: 4249

Name: RITTER2
Sterilizer ID: V2313228
01/08/2025 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.8 F 0.0
2:00 116.3 F 0.0
4:00 133.3 F 0.2
6:00 171.5 F 0.6
8:00 196.9 F 3.5
10:00 211.6 F 7.5
12:00 228.1 F 11.8
14:00 228.8 F 7.7
16:00 240.5 F 12.9
18:00 239.9 F 11.0
20:00 237.9 F 9.8
22:00 251.5 F 16.4
24:00 262.8 F 22.9
26:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.2 F 29.1
2:00 272.2 F 29.5
3:00 271.8 F 28.8
4:00 272.5 F 29.5

Min 271.4 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 233.6 F 0.0
10:00 231.0 F 0.0
15:00 223.7 F 0.0
20:00 223.0 F 0.0
25:00 220.0 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:17
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

01/08/2025 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____