

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2678

Name: RITTER 1  
Sterilizer ID: V2178198  
12/14/2022 03:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
2:00 203.3 F 4.2  
4:00 208.9 F 9.0  
6:00 225.4 F 7.0  
8:00 235.3 F 10.5  
10:00 235.3 F 7.9  
12:00 241.3 F 10.6  
14:00 247.7 F 13.9  
16:00 260.6 F 21.2  
18:00 272.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.2  
1:00 272.3 F 28.7  
2:00 271.8 F 28.6  
3:00 271.9 F 28.4  
4:00 272.3 F 29.0  
  
Min 271.5 F 28.1  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.3 F 0.0  
5:00 237.0 F 0.0  
10:00 238.1 F 0.0  
15:00 229.9 F 0.0  
20:00 227.5 F 0.0  
25:00 229.9 F 0.0  
30:00 228.2 F 0.0  
  
Drying Complete

FILLING: 1:53  
HEATING: 18:17  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:21  
CYCLE COMPLETE

12/14/2022 04:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_