
Midmark M11 - v1.0.5
Total Cycles: 2167

Name: RITTER2
Sterilizer ID: V2313228
03/22/2023 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 183.6 F 3.5
4:00 194.0 F 8.4
6:00 216.3 F 9.2
8:00 229.2 F 9.9
10:00 232.1 F 8.1
12:00 241.6 F 11.6
14:00 246.1 F 13.8
16:00 259.6 F 21.1
18:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 271.6 F 28.8
2:00 272.1 F 29.2
3:00 271.4 F 29.3
4:00 272.5 F 29.5

Min 271.0 F 28.7
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 217.6 F 0.0
10:00 236.5 F 0.0
15:00 241.0 F 0.0
20:00 238.3 F 0.0
25:00 226.8 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 18:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

03/22/2023 12:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____