

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5410

Name: RITTER 1  
Sterilizer ID: V2178198  
03/13/2025 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 69.6 F 0.3  
4:00 88.9 F 1.9  
6:00 127.2 F 4.5  
8:00 152.5 F 7.4  
10:00 173.9 F 11.0  
12:00 194.7 F 2.5  
14:00 209.7 F 4.9  
16:00 220.6 F 8.4  
18:00 230.7 F 12.3  
20:00 230.8 F 7.5  
22:00 242.1 F 12.5  
24:00 239.0 F 10.0  
26:00 251.6 F 16.1  
28:00 262.5 F 22.4  
30:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.5 F 28.1  
3:00 272.6 F 29.1  
4:00 271.9 F 28.4

Min 271.4 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
5:00 189.6 F 0.0  
10:00 185.2 F 0.0  
15:00 190.1 F 0.0  
20:00 189.6 F 0.0  
25:00 203.2 F 0.0  
30:00 217.3 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 30:25  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:08:36  
CYCLE COMPLETE

03/13/2025 09:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_