
Midmark M11 - v1.0.5
Total Cycles: 412

Name: RITTER2
Sterilizer ID: V2313228
10/04/2021 06:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 71.5 F 0.0
4:00 77.2 F 0.1
6:00 106.2 F 0.2
8:00 136.4 F 1.4
10:00 160.0 F 5.1
12:00 182.6 F 9.1
14:00 200.7 F 13.1
16:00 220.1 F 6.8
18:00 230.4 F 11.0
20:00 230.9 F 7.5
22:00 243.2 F 13.1
24:00 242.0 F 11.7
26:00 254.5 F 18.2
28:00 265.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.9 F 28.9
2:00 272.2 F 29.2
3:00 272.5 F 29.5
4:00 272.4 F 29.8

Min 271.6 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.2 F 0.0
5:00 190.6 F 0.0
10:00 190.4 F 0.0
15:00 190.4 F 0.0
20:00 186.9 F 0.0
25:00 182.3 F 0.0
30:00 177.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 29:44
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:07:45
CYCLE COMPLETE

10/04/2021 07:38 AM
ACCEPTED: X
REJECTED:
Operator: U01