
Midmark M11 - v1.0.0
Total Cycles: 3557

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.5 F 0.0
2:00 202.1 F 3.6
4:00 213.7 F 8.4
6:00 227.9 F 9.1
8:00 233.0 F 9.8
10:00 237.5 F 8.6
12:00 245.1 F 13.5
14:00 244.2 F 12.4
16:00 257.2 F 19.2
18:00 268.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.2 F 28.7
2:00 271.9 F 28.8
3:00 272.1 F 28.5
4:00 272.0 F 28.9

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 219.2 F 0.0
10:00 226.2 F 0.0
15:00 221.4 F 0.0
20:00 216.2 F 0.0
25:00 214.5 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

07/20/2023 01:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____