

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007164  
11:52 08/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:46  
STERILIZING 3:52  
136.0 C 220kPa 3:52  
136.4 C 214kPa 4:22  
136.0 C 217kPa 4:52  
136.2 C 220kPa 5:22  
136.3 C 217kPa 5:52  
CYCLE ABORTED 6:03  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 6:03  
-----

Digital Signature #  
2636E5732D19EFD8  
-----