
Midmark M11 - v1.0.0
Total Cycles: 2218

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 176.9 F 4.2
4:00 192.5 F 8.8
6:00 219.5 F 10.4
8:00 223.3 F 8.1
10:00 234.8 F 13.3
12:00 236.4 F 10.2
14:00 237.9 F 8.1
16:00 248.3 F 14.5
18:00 259.9 F 20.9
20:00 270.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.2 F 29.1
2:00 271.4 F 28.4
3:00 271.8 F 29.0
4:00 272.2 F 28.9

Min 271.1 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 217.1 F 0.0
10:00 231.5 F 0.0
15:00 228.7 F 0.0
20:00 224.8 F 0.0
25:00 227.9 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

08/09/2022 03:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____