
Midmark M11 - v1.0.5
Total Cycles: 107

Name: RITTER2
Sterilizer ID: V2313228
06/16/2021 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.5 F 0.0
2:00 141.9 F 1.8
4:00 155.7 F 4.6
6:00 181.0 F 8.1
8:00 197.6 F 11.7
10:00 211.5 F 5.3
12:00 222.7 F 9.5
14:00 232.3 F 11.1
16:00 238.3 F 11.4
18:00 238.4 F 10.1
20:00 249.8 F 15.8
22:00 260.4 F 21.8
24:00 269.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.3 F 30.2
2:00 272.6 F 29.9
3:00 271.7 F 29.8
4:00 272.8 F 30.3

Min 271.0 F 29.0
Max 273.5 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 228.2 F 0.0
10:00 203.6 F 0.0
15:00 185.2 F 0.0
20:00 180.2 F 0.0
25:00 164.8 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:02
CYCLE COMPLETE

06/16/2021 12:06 PM
ACCEPTED: X
REJECTED:
Operator: U01