
Midmark M11 - v1.0.0
Total Cycles: 5281

Name: RITTER 1
Sterilizer ID: V2178198
01/27/2025 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 185.9 F 3.9
4:00 191.5 F 8.3
6:00 207.5 F 13.3
8:00 220.7 F 7.5
10:00 234.3 F 12.6
12:00 235.0 F 9.3
14:00 239.5 F 10.0
16:00 246.3 F 13.7
18:00 258.1 F 20.1
20:00 268.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.5 F 29.1
2:00 272.0 F 29.2
3:00 271.5 F 28.7
4:00 271.2 F 28.9

Min 271.0 F 28.6
Max 272.3 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 210.6 F 0.0
10:00 196.9 F 0.0
15:00 196.0 F 0.0
20:00 200.4 F 0.0
25:00 211.4 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

01/27/2025 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____