

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1636

Name: RITTER2  
Sterilizer ID: V2313228  
10/21/2022 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
2:00 192.5 F 3.3  
4:00 199.4 F 7.8  
6:00 216.2 F 12.7  
8:00 225.3 F 7.7  
10:00 239.8 F 13.2  
12:00 239.8 F 11.0  
14:00 238.0 F 9.8  
16:00 252.0 F 16.7  
18:00 264.0 F 23.8  
20:00 271.2 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.3 F 29.8  
2:00 272.0 F 29.2  
3:00 271.2 F 29.0  
4:00 272.0 F 29.8  
  
Min 271.0 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 232.2 F 0.0  
10:00 234.1 F 0.0  
15:00 227.4 F 0.0  
20:00 230.2 F 0.0  
25:00 227.6 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

10/21/2022 11:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_