
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.3uS / 1.4ppm
CYCLE NUMBER 003676
8:16 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:07
STERILIZING 4:17
136.4 C 222kPa 4:17
136.2 C 220kPa 4:47
136.1 C 220kPa 5:17
136.3 C 223kPa 5:47
136.2 C 222kPa 6:17
136.2 C 220kPa 6:47
136.4 C 223kPa 7:17
136.0 C 222kPa 7:47
136.0 C 222kPa 7:47
VENTING 7:47
AIR DRYING 8:35
CYCLE COMPLETE 12:40

Digital Signature #
39341C330FF2E2A8
