
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007799
8:31 27/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:22
STERILIZING 3:10
136.1 C 220kPa 3:10
135.8 C 219kPa 3:40
136.4 C 221kPa 4:10
136.4 C 222kPa 4:40
136.4 C 221kPa 5:10
136.4 C 221kPa 5:40
136.2 C 219kPa 6:10
136.2 C 222kPa 6:40
136.2 C 222kPa 6:40
VENTING 6:40
CYCLE ABORTED 7:05
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:05

Digital Signature #
322EF56C534F8507
