

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2123

Name: RITTER2  
Sterilizer ID: V2313228  
03/08/2023 08:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.9 F 0.0  
2:00 80.4 F 0.0  
4:00 105.7 F 0.0  
6:00 110.7 F 0.0  
8:00 143.2 F 2.3  
10:00 177.4 F 5.6  
12:00 194.7 F 9.2  
14:00 213.1 F 13.0  
16:00 222.7 F 6.4  
18:00 234.0 F 10.6  
20:00 233.6 F 8.3  
22:00 241.3 F 12.1  
24:00 239.8 F 10.5  
26:00 252.1 F 16.7  
28:00 262.8 F 23.0  
30:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.2 F 29.1  
2:00 272.0 F 29.3  
3:00 271.8 F 28.9  
4:00 272.3 F 29.5

Min 271.1 F 28.6  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.0 F 0.0  
5:00 214.0 F 0.0  
10:00 238.5 F 0.0  
15:00 221.2 F 0.0  
20:00 226.3 F 0.0  
25:00 231.0 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:17  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

03/08/2023 09:16 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_