

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1507

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2022 10:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
2:00 192.7 F 3.8  
4:00 202.1 F 9.3  
6:00 223.5 F 7.1  
8:00 233.6 F 10.9  
10:00 235.0 F 9.1  
12:00 238.9 F 9.6  
14:00 250.3 F 16.0  
16:00 263.3 F 23.5  
18:00 271.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.6 F 29.6  
2:00 271.7 F 29.1  
3:00 272.7 F 29.7  
4:00 272.0 F 29.0  
  
Min 271.7 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 199.5 F 0.0  
10:00 214.6 F 0.0  
15:00 217.9 F 0.0  
20:00 221.4 F 0.0  
25:00 218.2 F 0.0  
30:00 217.4 F 0.0  
Drying Complete

FILLING: 2:58  
HEATING: 18:01  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:09  
CYCLE COMPLETE

09/20/2022 11:02 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_