
Midmark M11 - v1.0.5
Total Cycles: 2433

Name: RITTER2
Sterilizer ID: V2313228
05/30/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 188.6 F 2.6
4:00 189.8 F 7.2
6:00 210.8 F 12.8
8:00 225.3 F 8.3
10:00 235.8 F 11.7
12:00 240.3 F 11.9
14:00 239.7 F 10.7
16:00 253.2 F 17.5
18:00 264.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.7
2:00 271.8 F 29.1
3:00 271.3 F 29.2
4:00 272.5 F 29.8

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 226.7 F 0.0
10:00 228.6 F 0.0
15:00 229.4 F 0.0
20:00 230.1 F 0.0
25:00 228.0 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

05/30/2023 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____