
Midmark M11 - v1.0.0
Total Cycles: 4816

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2024 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.2 F 0.0
2:00 125.0 F 0.7
4:00 147.4 F 1.2
6:00 177.0 F 3.0
8:00 188.4 F 6.5
10:00 207.1 F 10.3
12:00 225.1 F 6.8
14:00 224.7 F 8.4
16:00 235.3 F 13.0
18:00 236.8 F 9.7
20:00 238.7 F 9.6
22:00 246.6 F 13.4
24:00 257.4 F 19.3
26:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 271.9 F 28.8
2:00 272.3 F 28.9
3:00 272.1 F 28.6
4:00 271.6 F 28.3

Min 271.2 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 215.0 F 0.0
10:00 222.0 F 0.0
15:00 225.8 F 0.0
20:00 227.1 F 0.0
25:00 230.6 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:36
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:16
CYCLE COMPLETE

08/14/2024 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____