
Midmark M11 - v1.0.0
Total Cycles: 2608

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 07:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 79.6 F 0.0
4:00 84.4 F 0.0
6:00 116.2 F 0.0
8:00 152.1 F 0.0
10:00 181.6 F 0.0
12:00 199.6 F 0.0
14:00 209.3 F 0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 211F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

11/28/2022 08:06 AM