
Midmark M11 - v1.0.5
Total Cycles: 1314

Name: RITTER2
Sterilizer ID: V2313228
07/18/2022 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.4 F 0.0
2:00 178.2 F 0.1
4:00 189.7 F 3.5
6:00 215.2 F 8.4
8:00 224.1 F 11.7
10:00 233.6 F 10.4
12:00 232.7 F 8.2
14:00 240.5 F 11.0
16:00 245.9 F 14.0
18:00 258.4 F 20.6
20:00 269.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.3 F 29.9
2:00 272.5 F 29.6
3:00 271.8 F 29.2
4:00 271.1 F 28.9

Min 271.0 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 230.5 F 0.0
10:00 239.9 F 0.0
15:00 224.2 F 0.0
20:00 220.2 F 0.0
25:00 218.2 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

07/18/2022 03:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____